

1 **Clinical Practice Guideline: Thermography**

2  
3 **Date of Implementation: February 9, 2006**

4  
5 **Product: Specialty**

6  
7  
8 **GUIDELINES**

9 The use of all forms of thermography is considered unproven and not medically necessary.

10  
11 **DESCRIPTION/BACKGROUND**

12 Thermography is a diagnostic procedure that measures surface temperature in various parts  
13 of the body. Thermography is used in various fields of medicine to help identify  
14 temperature abnormalities. There are four major types of thermography: liquid crystal;  
15 magnetic resonance (MR); thermocouple; and infrared thermography.

16  
17 Liquid crystal thermography involves the use of thermochromic liquid crystals that change  
18 their reflected color as a function of temperature. The simplest medical use of these is the  
19 forehead thermometer that is laid on the forehead to detect fever. They have also been used  
20 in the past to study skin diseases and lesions. This is an older technology which is no longer  
21 frequently used in medicine. Currently the most common use of liquid crystal  
22 thermography is in the electronics and technical arena.

23  
24 Magnetic resonance thermography uses the same technology as other magnetic resonance  
25 imaging (MRI) but incorporates the variable of temperature into the evaluation. The benefit  
26 of MR thermography is that it is very sensitive and can detect very small temperature  
27 changes, which are necessary for some uses.

28  
29 Thermocouples are temperature sensors that measure the change in voltage generated when  
30 any conductor is subjected to a thermal gradient (Seebeck effect). This voltage change is  
31 then transformed into a temperature reading. These devices are most commonly used in  
32 electronics and industrial contexts. In 1924, a chiropractor developed the neurocalometer  
33 – a handheld device with two thermocouples placed along the spine to measure temperature  
34 differences. The inventor theorized this device could detect subluxations through  
35 temperature differentials. The current version of the neurocalometer is the Nervoscope, a  
36 tool currently used by some chiropractors. Proponents believe viewing a patient’s skin  
37 thermal patterns provides valuable information on the normal and abnormal functioning of  
38 the underlying neurophysiology.

39  
40 Infrared thermography uses equipment sensitive only to infrared frequencies which convert  
41 the infrared radiation emitted from the skin surface into electrical impulses that may be  
42 visualized. The majority of these devices are large pieces of equipment, which need to be

1 used in a special climate-controlled room under controlled conditions. This type may be  
2 used for certain conditions including complex regional pain syndromes, carpal tunnel  
3 syndrome, disc herniation, and radiculopathy. There are also handheld infrared tools for  
4 which there is no reliable evidence of accuracy.

## 6 EVIDENCE REVIEW

7 The literature review focused on the use of thermography for musculoskeletal and related  
8 disorders. As such MRI thermography, which is used primarily for non-  
9 neuromusculoskeletal (NMS) conditions, was not considered in this evaluation. No  
10 randomized clinical trials for thermography in the peer review literature were found. Most  
11 studies of thermography have been performed comparing it with other diagnostic tests, the  
12 most common being clinical exam, computed tomography (CT), electromyography (EMG)  
13 and myelography. The bulk of the evidence regarding thermography is from  
14 methodological studies which provide little insight into the actual performance or utility of  
15 the instruments under investigation.

16  
17 Gulevich et al. (1997) found that infrared thermography was useful in the diagnosis of  
18 complex regional pain syndrome with a predictive validity of 90%. Herrick and Herrick  
19 (1987) and Ming et al. (2005) found that infrared thermography might be useful in  
20 diagnosing carpal tunnel syndrome and associated neuropathies. Ping and You (1993) and  
21 Zhang et al. (1999) found that infrared thermography may have utility as a diagnostic tool  
22 for patients with lumbar and cervical disc herniations respectively. Takahashi et al. (1994)  
23 concluded that thermography may be clinically useful in diagnosing radiculopathy based  
24 on the finding that the thermograms agreed with the findings of the clinical exam; however,  
25 they do not mention what type of thermography was utilized.

26  
27 Leclaire et al. (1996) found that thermography was not a useful diagnostic tool for low  
28 back pain but do not mention the type of thermography used; in addition, their paper had  
29 significant methodological flaws. So et al. (1989) found that thermographic findings were  
30 of little diagnostic value in the evaluation of lumbosacral radiculopathy, but they do not  
31 mention what type of thermography was used.

32  
33 Plaughter (1992) performed a systematic review of the literature on thermography for  
34 neuromusculoskeletal abnormalities of the spine, evaluating many different types of  
35 thermography. Plaughter concluded that full scan (non-handheld) infrared thermography  
36 was a sensitive diagnostic procedure for detecting spinal abnormalities such as disc  
37 protrusion. The review found that the evidence for liquid crystal thermography is weak and  
38 inconclusive and as such should not be recommended. For thermocouple and other  
39 handheld devices, the review found that there was no evidence supporting the use of these  
40 tools.

1 A literature review and meta-analysis by Hoffman et al. (1991) to determine the role of  
2 thermography for diagnosing lumbar radiculopathy produced no clear-cut results. The role  
3 of thermography remains unclear. Rigorous clinical research is required to establish its  
4 diagnostic accuracy and clinical utility. Thermography cannot be recommended currently  
5 for routine clinical use in evaluating low-back pain.

6  
7 DiBenedetto et al. (2002) reported that thermograms of injured feet show areas of increased  
8 heat, but excessive weight-bearing pressure on feet, new shoes, or boots also cause  
9 increased infrared emission even without discomfort. Differentiation remains difficult;  
10 thermography may detect early injury; however, it does not reveal exact diagnoses.

11  
12 To evaluate the perceived status of thermography in the diagnosis of musculoskeletal  
13 disorders, Awerbuch (1991) used the following data sources: medical and legal journals  
14 published from 1956 onward; report of the United States Office of Health Technology  
15 Assessment and personal communication with the author of that report. The selection of  
16 studies was confined to application of thermography to musculoskeletal and neurological  
17 medicine. Awerbuch concluded that little evidence exists of any application of  
18 thermography in which it is unequivocally superior to conventional diagnostic imaging  
19 methods.

20  
21 Meeker and Gahlinger (1986) provide a review and summary of research and a comparison  
22 with myelography, computerized tomography, electromyography, and clinical and surgical  
23 findings in cases of presumed musculoskeletal pain syndromes. The importance of  
24 diagnostic sensitivity, specificity, positive and negative predictive value, and accuracy  
25 (validity) are discussed. In general, the literature reports high sensitivity and negative  
26 predictive value, but lower specificity and positive predictive value. They examine the  
27 implications of these findings in regard to clinical case management, with emphasis on  
28 potential usefulness to chiropractors. There remain a number of questions about the  
29 scientific validity of this research.

30  
31 A systematic review by Triano et al. (2013) found that there was good evidence for the  
32 reliability of thermography in identifying lower limb sciatica. However, the review  
33 concluded that there was no good evidence to support the use of thermography to measure  
34 paraspinal temperatures for the purpose of localizing the site of care. Sanchis- Sánchez et  
35 al. (2014) completed a systematic review and meta-analysis on infrared thermal imaging  
36 in the diagnosis of musculoskeletal injuries. The authors concluded there is a lack of  
37 support for the usefulness of infrared thermal imaging in musculoskeletal injury diagnosis.  
38 Dibai-Filho and Guirro (2015) did a critical review of the literature on the evaluation of  
39 myofascial trigger points using infrared thermography. The authors concluded that  
40 currently, there are few studies evaluating the accuracy and reliability of infrared  
41 thermography for the diagnosis and assessment of myofascial trigger points.

1 Schiavon et al. (2021) aimed at assessing infrared thermography potential and limitations  
2 in these pathologies in a systematic review. Of 718 screened articles, 32 were found to be  
3 eligible for inclusion, for a total of 2,094 patients. Nine studies reported the application to  
4 osteoarthritis, 21 to rheumatic diseases, 2 on both. The publication trend showed an  
5 increasing interest in the last decade. Seven studies investigated the correlation of  
6 temperature changes with osteoarthritis, 16 with rheumatic diseases, and 2 with both,  
7 whereas 2 focused on the pre-post evaluation to investigate treatment results in patients  
8 with osteoarthritis and 5 in patients with rheumatic diseases. A correlation was shown  
9 between thermal findings and disease presence and stage, as well as the clinical assessment  
10 of disease activity and response to treatment, supporting infrared thermography role in the  
11 study and management of rheumatic diseases and osteoarthritis. Authors conclude that this  
12 systematic literature review showed an increasing interest in this technology, with several  
13 applications in different joints affected by inflammatory and degenerative pathologies.  
14 Infrared thermography proved to be a simple, accurate, noninvasive, and radiation-free  
15 method, which could be used in addition to the currently available tools for screening,  
16 diagnosis, monitoring of disease progression, and response to medical treatment.

17  
18 Albuquerque et al. (2021) evaluated the role of infrared thermography as a helpful outcome  
19 measure tool in subjects with back and neck syndromes in a systematic review. From these,  
20 268 duplicates were removed, and only 16 were in line with the aim of this review.  
21 Ultimately, only seven precisely fulfilled the inclusion and exclusion criteria and were  
22 included in the review. According to the articles reviewed, thermography seems to give an  
23 objective notion of change in inflammatory activity, which can corroborate the usefulness  
24 of treatment or the improvement/worsening of the patient's symptoms. The overall quality  
25 of research was uneven in the study design, endpoint measures, and sample characteristics.  
26 Authors concluded that the number of high-quality studies of the role of infrared  
27 thermography in patients with back and neck syndromes remains limited. More than a  
28 diagnostic tool, thermography can be an objective tool for monitoring the effectiveness of  
29 treatment by identifying deviations from a healthy state.

30  
31 Park et al. (2021) investigated the thermographic findings of carpal tunnel syndrome  
32 (CTS). They enrolled 304 hands with electrodiagnostically identified CTS and 88 control  
33 hands. CTS hands were assigned to duration groups (D1, < 3 months; D2, 3–6 months; D3,  
34 6–12 months; D4, ≥ 12 months) and severity groups (S1, very mild; S2, mild; S3, moderate;  
35 S4, severe). The temperature difference between the median and ulnar nerve territories  
36 ( $\Delta M-U$  territories) decreased as CTS duration and severity increased. Significant  
37 differences in  $\Delta M-U$  territories between the D1 and D3, D1 and D4, D2 and D4, and S1  
38 and S4 groups were observed. Thermal anisometry increased as CTS duration and severity  
39 increased. Significant differences in thermal anisometry between the D1 and D4 as well as  
40 the D2 and D4 groups were noted. Thermal anisometry was higher in the S4 group than in  
41 the S1, S2, and S3 groups. As CTS progresses, skin temperature tends to decrease and

1 thermal variation tends to increase in the median nerve-innervated area. Thermographic  
2 findings reflect the physiological changes of the entrapped median nerve.

3  
4 Dias de Lacerda et al. (2022) systematically reviewed the literature on diagnosis of  
5 tendinopathy using infrared thermography (IT). Seven studies were included in the meta-  
6 analyses, which showed that the IT has an overall sensitivity of 72% and specificity of  
7 95%. The IT showed adequate accuracy to detect tendon injuries, with high specificity in  
8 the evaluation of lateral epicondylitis and shoulder tendinopathy.

### 9 10 **References**

11 Albuquerque NF, Lopes BS. Musculoskeletal applications of infrared thermography on  
12 back and neck syndromes: a systematic review. *Eur J Phys Rehabil Med.*  
13 2021;57(3):386-396

14  
15 Awerbuch MS. Thermography--its current diagnostic status in musculoskeletal  
16 medicine. *Med J Aust.* 1991;154(7):441-444. doi:10.5694/j.1326-  
17 5377.1991.tb121171.x

18  
19 BenEliyah, D. J. Protocols and standards for thermography imaging—Part I. *Dynamic*  
20 *Chiropractic.* 1992;10(22). Retrieved March 26, 2024, from  
21 <http://www.dynamicchiropractic.com/mpacms/dc/article.php?id=43558>

22  
23 BenEliyah, D. J. Thermography in clinical practice: The rebuttal. *Dynamic Chiropractic.*  
24 1993;11(14). Retrieved March 26, 2024, from  
25 <http://www.dynamicchiropractic.com/mpacms/dc/article.php?id=42385>

26  
27 Centers for Medicare and Medicaid Services. National Coverage Determination (NCD) for  
28 Thermography (220.11). Retrieved on March 26, 2024 from  
29 [https://www.cms.gov/medicare-coverage-database/details/ncd-](https://www.cms.gov/medicare-coverage-database/details/ncd-details.aspx?NCDId=164&ncdver=1&bc=AAAAGAAAAAAA&)  
30 [details.aspx?NCDId=164&ncdver=1&bc=AAAAGAAAAAAA&](https://www.cms.gov/medicare-coverage-database/details/ncd-details.aspx?NCDId=164&ncdver=1&bc=AAAAGAAAAAAA&)

31  
32 Dias de Lacerda AP, Rodrigues de Andrade P, Kamonseki DH, et al. Accuracy of infrared  
33 thermography in detecting tendinopathy: A systematic review with meta-analysis. *Phys*  
34 *Ther Sport.* 2022;58:117-125. doi:10.1016/j.ptsp.2022.10.005

35  
36 Dibai-Filho AV, Guirro RRJ. Evaluation of myofascial trigger points using infrared  
37 thermography: a critical review of the literature. *J Manipulative Physiol Ther.*  
38 2015;38(1):86-92. doi:10.1016/j.jmpt.2014.10.010

39  
40 DiBenedetto M, Yoshida M, Sharp M, Jones B. Foot evaluation by infrared imaging. *Mil*  
41 *Med.* 2002;167(5):384-392

- 1 Edeiken J, Shaber G. Thermography: a reevaluation. *Skeletal Radiol.* 1986;15(7):545-548.  
2 doi:10.1007/BF00361052  
3
- 4 Fitzgerald A, Berentson-Shaw J. Thermography as a screening and diagnostic tool: a  
5 systematic review. *N Z Med J.* 2012;125(1351):80-91. Published 2012 Mar 9  
6
- 7 Gulevich SJ, Conwell TD, Lane J, et al. Stress infrared telethermography is useful in the  
8 diagnosis of complex regional pain syndrome, type I (formerly reflex sympathetic  
9 dystrophy). *Clin J Pain.* 1997;13(1):50-59. doi:10.1097/00002508-199703000-00008  
10
- 11 Han, S. S., Jung, C. H., Lee, S. C., Jung, H. J., & Kim, Y. H. Does skin temperature  
12 difference as measured by infrared thermography within 6 months of acute herpes  
13 zoster infection correlate with pain level? *Skin research and technology: official journal*  
14 *of International Society for Bioengineering and the Skin (ISBS) [and] International*  
15 *Society for Digital Imaging of Skin (ISDIS) [and] International Society for Skin*  
16 *Imaging (ISSI).* 2010; 16(2), 198–201. [https://doi.org/10.1111/j.1600-](https://doi.org/10.1111/j.1600-0846.2009.00417.x)  
17 [0846.2009.00417.x](https://doi.org/10.1111/j.1600-0846.2009.00417.x)  
18
- 19 Herrick, R. T. and S. K. Herrick. Thermography in the detection of carpal tunnel syndrome  
20 and other compressive neuropathies. *Journal of Hand Surgery [Am].* 1987; 12(5 Pt 2):  
21 943-9.  
22
- 23 Hoffman RM, Kent DL, Deyo RA. Diagnostic accuracy and clinical utility of  
24 thermography for lumbar radiculopathy. A meta-analysis. *Spine (Phila Pa 1976).*  
25 1991;16(6):623-628. doi:10.1097/00007632-199106000-00005  
26
- 27 International Academy of Clinical Thermology. Breast. Retrieved on April 2, 2024 from  
28 <http://www.iact-org.org>  
29
- 30 Jones B, Hassan I, Tsuyuki RT, Dos Santos MF, Russell AS, Yacyshyn E. Hot joints: myth  
31 or reality? A thermographic joint assessment of inflammatory arthritis patients. *Clin*  
32 *Rheumatol.* 2018;37(9):2567-2571. doi:10.1007/s10067-018-4108-0  
33
- 34 Knowitz, K. Thermography: A new perspective on an old test. *Dynamic Chiropractic.*  
35 2000;18(13). Retrieved March 26, 2024 from  
36 <http://www.dynamicchiropractic.com/mpacms/dc/article.php?id=31740>  
37
- 38 Leclaire R, Esdaile JM, Jéquier JC, Hanley JA, Rossignol M, Bourdouxhe M. Diagnostic  
39 accuracy of technologies used in low back pain assessment. Thermography, triaxial  
40 dynamometry, spinoscopy, and clinical examination. *Spine (Phila Pa 1976).*  
41 1996;21(11):1325-1331. doi:10.1097/00007632-199606010-00009

- 1 Meeker WC, Gahlinger PM. Neuromusculoskeletal thermography: a valuable diagnostic  
2 tool?. *J Manipulative Physiol Ther.* 1986;9(4):257-266  
3
- 4 Ming Z, Zaproudina N, Siivola J, Nousiainen U, Pietikainen S. Sympathetic pathology  
5 evidenced by hand thermal anomalies in carpal tunnel syndrome. *Pathophysiology.*  
6 2005;12(2):137-141. doi:10.1016/j.pathophys.2005.05.002  
7
- 8 Omranipour R, Kazemian A, Alipour S, et al. Comparison of the Accuracy of  
9 Thermography and Mammography in the Detection of Breast Cancer. *Breast Care*  
10 (Basel). 2016;11(4):260-264. doi:10.1159/000448347  
11
- 12 Park D, Kim BH, Lee SE, et al. Application of digital infrared thermography for carpal  
13 tunnel syndrome evaluation. *Sci Rep.* 2021;11(1):21963. Published 2021 Nov 9  
14
- 15 Ping Z, You FT. Correlation study on infrared thermography and nerve root signs in lumbar  
16 intervertebral disk herniation patient: a short report [published correction appears in *J*  
17 *Manipulative Physiol Ther* 1993 Oct;16(8):560]. *J Manipulative Physiol Ther.*  
18 1993;16(3):150-154  
19
- 20 Plaugher G. Skin temperature assessment for neuromusculoskeletal abnormalities of the  
21 spinal column. *J Manipulative Physiol Ther.* 1992;15(6):365-381  
22
- 23 Robinson V, Brosseau L, Casimiro L, et al. Thermotherapy for treating rheumatoid  
24 arthritis. *Cochrane Database Syst Rev.* 2002;(1):CD002826.  
25 doi:10.1002/14651858.CD002826  
26
- 27 Sanchis-Sánchez E, Vergara-Hernández C, Cibrián RM, Salvador R, Sanchis E, Codoñer-  
28 Franch P. Infrared thermal imaging in the diagnosis of musculoskeletal injuries: a  
29 systematic review and meta-analysis. *AJR Am J Roentgenol.* 2014;203(4):875-882.  
30 doi:10.2214/AJR.13.11716  
31
- 32 Schiavon G, Capone G, Frize M, Zaffagnini S, Candrian C, Filardo G. Infrared  
33 Thermography for the Evaluation of Inflammatory and Degenerative Joint Diseases: A  
34 Systematic Review. *Cartilage.* 2021;13(2\_suppl):1790S-1801S  
35
- 36 Story, R. J. (1994, January 14). Chiropractic thermography: Objectifying subluxation.  
37 *Dynamic Chiropractic*, 12(2). Retrieved March 26, 2024, from  
38 <http://www.dynamicchiropractic.com/mpacms/dc/article.php?id=41044>  
39
- 40 So YT, Aminoff MJ, Olney RK. The role of thermography in the evaluation of lumbosacral  
41 radiculopathy. *Neurology.* 1989;39(9):1154-1158. doi:10.1212/wnl.39.9.1154

- 1 Takahashi Y, Takahashi K, Moriya H. Thermal deficit in lumbar radiculopathy.  
2 Correlations with pain and neurologic signs and its value for assessing symptomatic  
3 severity. *Spine (Phila Pa 1976)*. 1994;19(21):2443-2450  
4
- 5 Triano JJ, Budgell B, Bagnulo A, et al. Review of methods used by chiropractors to  
6 determine the site for applying manipulation. *Chiropr Man Therap*. 2013;21(1):36.  
7 Published 2013 Oct 21. doi:10.1186/2045-709X-21-36  
8
- 9 Vreugdenburg TD, Willis CD, Mundy L, Hiller JE. A systematic review of elastography,  
10 electrical impedance scanning, and digital infrared thermography for breast cancer  
11 screening and diagnosis. *Breast Cancer Res Treat*. 2013;137(3):665-676.  
12 doi:10.1007/s10549-012-2393-x  
13
- 14 Woźniak K, Szyszka-Sommerfeld L, Trybek G, Piątkowska D. Assessment of the  
15 Sensitivity, Specificity, and Accuracy of Thermography in Identifying Patients with  
16 TMD. *Med Sci Monit*. 2015;21:1485-1493. Published 2015 May 23.  
17 doi:10.12659/MSM.893863  
18
- 19 Zhang HY, Kim YS, Cho YE. Thermatomal changes in cervical disc herniations. *Yonsei*  
20 *Med J*. 1999;40(5):401-412. doi:10.3349/ymj.1999.40.5.401